

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
 Data File : VN055363.D
 Acq On : 2 May 2019 21:46
 Operator : JC/SP
 Sample : VN0502MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0502MBL02

Quant Time: May 03 05:28:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	394740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	673090	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	609705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	178283	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	271484	52.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.14%	
35) Dibromofluoromethane	7.59	113	214806	48.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.24%	
50) Toluene-d8	10.09	98	824665	51.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.38%	
62) 4-Bromofluorobenzene	12.40	95	261822	52.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
Data File : VN055363.D
Acq On : 2 May 2019 21:46
Operator : JC/SP
Sample : VN0502MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0502MBL02

Quant Time: May 03 05:28:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

