

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
 Data File : VN055368.D
 Acq On : 2 May 2019 23:52
 Operator : JC/SP
 Sample : K2658-04 10X
 Misc : 6.25g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-SB-E6(17-20)

Quant Time: May 03 05:56:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	402859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	698555	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	635759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	261866	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	277416	52.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.28%	
35) Dibromofluoromethane	7.59	113	220252	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
50) Toluene-d8	10.09	98	873761	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	
62) 4-Bromofluorobenzene	12.40	95	413434	79.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	159.80%	
Target Compounds						
73) Isopropylbenzene	12.25	105	775235	36.342	ug/l	100
78) n-propylbenzene	12.59	91	2331522	86.677	ug/l	100
83) tert-Butylbenzene	12.99	119	159859	8.591	ug/l	89
84) 1,2,4-Trimethylbenzene	13.04	105	92765	5.052	ug/l	99
85) sec-Butylbenzene	13.17	105	1201771	53.858	ug/l #	65
89) n-Butylbenzene	13.61	91	621835	47.716	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
Data File : VN055368.D
Acq On : 2 May 2019 23:52
Operator : JC/SP
Sample : K2658-04 10X
Misc : 6.25g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-E6(17-20)

Quant Time: May 03 05:56:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

