

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
 Data File : VN055372.D
 Acq On : 3 May 2019 2:22
 Operator : JC/SP
 Sample : K2658-13 10X
 Misc : 6.29g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP1-SB-FD-01

Quant Time: May 03 10:01:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 30 07:16:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	450242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	757998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	689608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	278060	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	295783	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
35) Dibromofluoromethane	7.59	113	235975	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	10.09	98	942057	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
62) 4-Bromofluorobenzene	12.40	95	416619	74.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	148.40%	
Target Compounds						
73) Isopropylbenzene	12.25	105	596437	25.402	ug/l	99
78) n-propylbenzene	12.59	91	1707885	59.795	ug/l	100
83) tert-Butylbenzene	12.99	119	123978	5.441	ug/l	89
84) 1,2,4-Trimethylbenzene	13.04	105	62308	3.196	ug/l	99
85) sec-Butylbenzene	13.17	105	935324	39.476	ug/l #	67
89) n-Butylbenzene	13.61	91	484864	35.039	ug/l	89
95) Naphthalene	15.13	128	9997	4.536	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
Data File : VN055372.D
Acq On : 3 May 2019 2:22
Operator : JC/SP
Sample : K2658-13 10X
Misc : 6.29g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-FD-01

Quant Time: May 03 10:01:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

