


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN050219\
Data File : VN055361.D
Acq On : 2 May 2019 20:31
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Title : SW846 8260
Last Update : Tue Apr 30 07:16:52 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.0	10816	PASS
75	95	30	60	54.3	26712	PASS
95	95	100	100	100.0	49188	PASS
96	95	5	9	6.7	3300	PASS
173	174	0.00	2	1.8	726	PASS
174	95	50	100	80.2	39453	PASS
175	174	5	9	8.2	3220	PASS
176	174	95	101	99.5	39274	PASS
177	176	5	9	5.9	2333	PASS