

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
 Data File : VN072295.D
 Acq On : 02 May 2022 14:47
 Operator : JC\MD
 Sample : N2619-24
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SOIL-EQUIPMENT-RINSE

Quant Time: May 03 03:12:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	189807	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	316896	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	268898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	81575	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	140253	48.131	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.260%
35) Dibromofluoromethane	8.016	113	89557	44.153	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	88.300%
50) Toluene-d8	10.439	98	318973	40.021	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	80.040%
62) 4-Bromofluorobenzene	12.727	95	97838	36.984	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	73.960%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	8494	6.761	ug/l	# 76
25) 2-Butanone	7.339	43	5742	2.870	ug/l	# 84
59) 2-Hexanone	11.080	43	2636	1.011	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
Data File : VN072295.D
Acq On : 02 May 2022 14:47
Operator : JC\MD
Sample : N2619-24
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SOIL-EQUIPMENT-RINSE

Quant Time: May 03 03:12:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

