

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
 Data File : VN072296.D
 Acq On : 02 May 2022 15:11
 Operator : JC\MD
 Sample : N2619-26
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: May 03 03:12:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	188694	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	318826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	275555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	82277	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	179451	61.946	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	123.900%
35) Dibromofluoromethane	8.016	113	114440	56.080	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	112.160%
50) Toluene-d8	10.439	98	404218	50.410	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.820%
62) 4-Bromofluorobenzene	12.727	95	122251	45.933	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
Data File : VN072296.D
Acq On : 02 May 2022 15:11
Operator : JC\MD
Sample : N2619-26
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: May 03 03:12:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

