

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
 Data File : VN072297.D
 Acq On : 02 May 2022 15:35
 Operator : JC\MD
 Sample : N2624-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N47992

Quant Time: May 03 03:12:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	192899	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	315604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	277643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	92040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	168106	56.765	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 113.540%	
35) Dibromofluoromethane	8.016	113	108457	53.690	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.380%	
50) Toluene-d8	10.439	98	386254	48.662	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 97.320%	
62) 4-Bromofluorobenzene	12.727	95	125169	47.509	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.020%	
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	9567	7.493	ug/l	# 78
25) 2-Butanone	7.328	43	4597700	2261.110	ug/l	95
67) Ethyl Benzene	11.839	91	25245	2.580	ug/l	92
68) m/p-Xylenes	11.945	106	6750	1.858	ug/l	91
69) o-Xylene	12.274	106	2360	0.670	ug/l	96
84) 1,2,4-Trimethylbenzene	13.362	105	3903	0.596	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
Data File : VN072297.D
Acq On : 02 May 2022 15:35
Operator : JC\MD
Sample : N2624-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N47992

Quant Time: May 03 03:12:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

