

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
 Data File : VN072298.D
 Acq On : 02 May 2022 16:15
 Operator : JC\MD
 Sample : N2624-01DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N47992DL

Quant Time: May 03 03:12:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	192661	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	310703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	269925	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	86042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	151112	51.090	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.180%
35) Dibromofluoromethane	8.022	113	96406	48.478	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.960%
50) Toluene-d8	10.439	98	341679	43.725	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	87.440%
62) 4-Bromofluorobenzene	12.727	95	104840	40.421	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	80.840%
Target Compounds						
25) 2-Butanone	7.328	43	406631	200.225	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050222\
Data File : VN072298.D
Acq On : 02 May 2022 16:15
Operator : JC\MD
Sample : N2624-01DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N47992DL

Quant Time: May 03 03:12:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

