

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
 Data File : VN081973.D
 Acq On : 02 May 2024 11:51
 Operator : JC\MD
 Sample : P2343-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240418002-09-TRIP-BLANK

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 05/03/2024
 Supervised By :Semsettin Yesilyurt 05/03/2024

Quant Time: May 03 00:10:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	32239	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	209838	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	174360	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	88246	32.693	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	= 108.967%		
60) 4-Bromofluorobenzene	12.853	95	78857	27.176	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	= 90.600%		
63) Toluene-d8	10.571	98	259320	31.317	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	= 104.400%		
Target Compounds						Qvalue
25) Chloroform	7.971	83	21922m	5.630	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
Data File : VN081973.D
Acq On : 02 May 2024 11:51
Operator : JC\MD
Sample : P2343-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240418002-09-TRIP-BLANK

Quant Time: May 03 00:10:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 05/03/2024
Supervised By :Semsettin Yesilyurt 05/03/2024

