

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
 Data File : VN081976.D
 Acq On : 02 May 2024 13:02
 Operator : JC\MD
 Sample : P2345-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240430044-02

Quant Time: May 03 00:11:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	35892	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	208201	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	172641	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	89127	29.659	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.867%
60) 4-Bromofluorobenzene	12.847	95	81088	28.223	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.067%
63) Toluene-d8	10.565	98	254562	31.049	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.500%
Target Compounds					Qvalue	
62) Toluene	10.630	91	2398	0.247	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
Data File : VN081976.D
Acq On : 02 May 2024 13:02
Operator : JC\MD
Sample : P2345-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240430044-02

Quant Time: May 03 00:11:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

