

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
 Data File : VN081980.D
 Acq On : 02 May 2024 16:30
 Operator : JC\MD
 Sample : P2361-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240501008-01

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 05/03/2024
 Supervised By :Semsettin Yesilyurt 05/03/2024

Quant Time: May 03 00:12:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	34022	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	202967	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	171985	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	88907	31.212	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 104.033%			
60) 4-Bromofluorobenzene	12.847	95	79028	27.611	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 92.033%			
63) Toluene-d8	10.571	98	251928	30.845	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 102.800%			
Target Compounds						Qvalue
15) Acetone	4.430	58	1879m	7.066	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
Data File : VN081980.D
Acq On : 02 May 2024 16:30
Operator : JC\MD
Sample : P2361-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240501008-01

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 05/03/2024
Supervised By :Semsettin Yesilyurt 05/03/2024

Quant Time: May 03 00:12:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Apr 27 04:37:27 2024
Response via : Initial Calibration

