

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
 Data File : VN081981.D
 Acq On : 02 May 2024 17:13
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: May 03 00:12:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	79	0.00
2 M	Dichlorodifluoromethane	1.703	1.651	3.1	77	0.00
3 M	Chloromethane	1.975	1.952	1.2	81	0.00
4 M	Vinyl Chloride	1.866	1.832	1.8	81	0.00
5 M	Bromomethane	0.974	0.819	15.9	67	0.00
6 M	Chloroethane	1.056	1.093	-3.5	81	0.00
7 M	Trichlorofluoromethane	2.730	2.585	5.3	75	0.00
8 T	Diethyl Ether	1.322	1.297	1.9	77	0.00
9	1,1,2-Trichlorotrifluoroeth	1.872	1.782	4.8	74	0.02
10 M	1,1-Dichloroethene	1.831	1.810	1.1	80	0.00
11	Methyl Iodide	1.673	1.622	3.0	89	-0.01
12	Methyl Acetate	2.293	2.572	-12.2	91	0.00
13 M	Acrolein	0.487	0.361	25.9#	62	0.00
14 M	Acrylonitrile	1.046	1.108	-5.9	84	0.00
15 M	Acetone	0.234	0.239	-2.1	79	0.00
16 M	Carbon Disulfide	5.073	4.724	6.9	77	0.00
17	Allyl chloride	3.215	3.168	1.5	76	0.00
18 M	Methylene Chloride	2.083	2.065	0.9	82	-0.01
19 M	trans-1,2-Dichloroethene	1.906	1.863	2.3	82	0.00
20 T	Diisopropyl ether	7.112	7.394	-4.0	83	0.00
21 M	1,1-Dichloroethane	3.766	3.883	-3.1	81	0.00
22 M	cis-1,2-Dichloroethene	2.322	2.317	0.2	82	0.00
23 M	tert-Butyl Alcohol	0.431	0.405	6.0	71	0.01
24 M	Methyl tert-Butyl Ether	6.601	6.631	-0.5	78	0.00
25 M	Chloroform	3.623	3.772	-4.1	82	0.00
26	Cyclohexane	3.257	3.219	1.2	81	0.00
27 s	1,2-Dichloroethane-d4	2.512	2.468	1.8	79	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
29	1,1-Dichloropropene	0.447	0.419	6.3	79	0.00
30 M	2-Butanone	0.230	0.228	0.9	81	0.00
31	2,2-Dichloropropane	0.458	0.520	-13.5	97	-0.01
32 M	1,1,1-Trichloroethane	0.514	0.501	2.5	81	0.00
33 M	Carbon Tetrachloride	0.417	0.396	5.0	78	0.00
34 M	Benzene	1.425	1.405	1.4	81	0.00
35	Methacrylonitrile	0.291	0.272	6.5	79	0.00
36 M	1,2-Dichloroethane	0.478	0.439	8.2	77	0.00
37 M	Trichloroethene	0.349	0.295	15.5	71	0.00
38	Methylcyclohexane	0.540	0.518	4.1	80	0.00
39 M	1,2-Dichloropropane	0.366	0.367	-0.3	86	0.00
40	Dibromomethane	0.235	0.233	0.9	84	0.00
41 M	Bromodichloromethane	0.502	0.466	7.2	79	0.00
42 M	Vinyl Acetate	0.980	0.955	2.6	82	0.00
43	Ethyl Acetate	0.502	0.504	-0.4	85	0.01
44	Isopropyl Acetate	0.891	0.865	2.9	81	0.00
45 T	1,4-Dioxane	0.007	0.007	0.0	84	0.00
46	Methyl methacrylate	0.407	0.399	2.0	82	0.00
47	n-amyl Acetate	0.804	0.773	3.9	82	0.00
48 M	t-1,3-Dichloropropene	0.544	0.518	4.8	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050224\
 Data File : VN081981.D
 Acq On : 02 May 2024 17:13
 Operator : JC\MD
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: May 03 00:12:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N042624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Apr 27 04:37:27 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.576	0.550	4.5	80	0.00
50 M	1,1,2-Trichloroethane	0.319	0.322	-0.9	84	0.00
51	Ethyl methacrylate	0.601	0.591	1.7	83	0.00
52	1,3-Dichloropropane	0.585	0.570	2.6	82	0.00
53 M	Dibromochloromethane	0.347	0.322	7.2	81	0.00
54 M	1,2-Dibromoethane	0.318	0.318	0.0	83	0.00
55 M	2-Chloroethyl vinyl ether	0.301	0.319	-6.0	88	0.00
56 M	Bromoform	0.211	0.178	15.6	71	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
58 M	4-Methyl-2-Pentanone	0.583	0.579	0.7	80	0.00
59 M	2-Hexanone	0.432	0.435	-0.7	81	0.00
60 S	4-Bromofluorobenzene	0.499	0.492	1.4	83	0.00
61 M	Tetrachloroethene	0.333	0.298	10.5	71	0.00
62 M	Toluene	1.688	1.692	-0.2	81	0.00
63 S	Toluene-d8	1.425	1.462	-2.6	83	0.00
64 M	Chlorobenzene	1.032	1.012	1.9	79	0.00
65	1,1,1,2-Tetrachloroethane	0.338	0.321	5.0	76	0.00
66 M	Ethyl Benzene	1.850	1.828	1.2	78	0.00
67 M	m/p-Xylenes	0.686	0.685	0.1	80	0.00
68 M	o-Xylene	0.679	0.674	0.7	79	0.00
69 M	Styrene	1.139	1.124	1.3	81	0.00
70	Isopropylbenzene	1.724	1.682	2.4	78	0.00
71 M	1,1,2,2-Tetrachloroethane	0.496	0.523	-5.4	83	0.00
72	1,2,3-Trichloropropane	0.464	0.471	-1.5	84	0.00
73	Bromobenzene	0.369	0.347	6.0	78	0.00
74	n-propylbenzene	2.034	2.080	-2.3	82	0.00
75	2-Chlorotoluene	1.258	1.204	4.3	77	0.00
76	1,3,5-Trimethylbenzene	1.395	1.379	1.1	79	0.00
77	t-1,4-Dichloro-2-butene	0.206	0.221	-7.3	89	0.00
78	4-Chlorotoluene	1.233	1.214	1.5	81	0.00
79	tert-butylbenzene	1.201	1.164	3.1	77	0.00
80	1,2,4-Trimethylbenzene	1.401	1.363	2.7	79	0.00
81	sec-Butylbenzene	1.684	1.726	-2.5	83	0.00
82	p-Isopropyltoluene	1.332	1.329	0.2	79	0.00
83 M	1,3-Dichlorobenzene	0.664	0.644	3.0	77	0.00
84 M	1,4-Dichlorobenzene	0.668	0.652	2.4	79	0.00
85	n-Butylbenzene	1.216	1.232	-1.3	84	0.00
86 T	Hexachloroethane	0.266	0.242	9.0	73	0.00
87 M	1,2-Dichlorobenzene	0.662	0.642	3.0	79	0.00
88	1,2-Dibromo-3-Chloropropane	0.111	0.102	8.1	71	0.00
89	1,2,4-Trichlorobenzene	0.364	0.323	11.3	73	0.00
90	Hexachlorobutadiene	0.135	0.108	20.0	69	0.00
91 M	Naphthalene	1.385	1.275	7.9	74	0.00
92	1,2,3-Trichlorobenzene	0.364	0.323	11.3	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0