

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050225\
 Data File : VN086461.D
 Acq On : 02 May 2025 14:15
 Operator : JC\MD
 Sample : PB167816ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167816ZHE#06

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/05/2025
 Supervised By : Semsettin Yesilyurt 05/05/2025

Quant Time: May 02 23:16:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	292125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114400	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	110877	49.995	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.980%		
35) Dibromofluoromethane	8.159	113	73784	54.421	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	108.840%		
50) Toluene-d8	10.565	98	383001	52.857	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	105.720%		
62) 4-Bromofluorobenzene	12.847	95	133527	50.523	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	101.040%		
Target Compounds						
16) Acetone	4.436	43	26760	29.875	ug/l	98
20) Methylene Chloride	5.283	84	9083	4.430	ug/l #	90
43) Isopropyl Acetate	8.688	43	33284m	4.725	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050225\
Data File : VN086461.D
Acq On : 02 May 2025 14:15
Operator : JC\MD
Sample : PB167816ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167816ZHE#06

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/05/2025
Supervised By :Semsettin Yesilyurt 05/05/2025

Quant Time: May 02 23:16:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration

