

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050225\
 Data File : VN086462.D
 Acq On : 02 May 2025 14:40
 Operator : JC\MD
 Sample : PB167816ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167816ZHE#07

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/05/2025
 Supervised By : Semsettin Yesilyurt 05/05/2025

Quant Time: May 02 23:16:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 16 04:19:23 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	154085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293175	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	277547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112459	50.331	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.660%			
35) Dibromofluoromethane	8.165	113	73051	53.688	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.380%			
50) Toluene-d8	10.565	98	372970	51.288	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.580%			
62) 4-Bromofluorobenzene	12.847	95	127868	48.208	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.420%			
Target Compounds						Qvalue
16) Acetone	4.424	43	25215	27.562	ug/l	94
20) Methylene Chloride	5.283	84	8464	4.098	ug/l	90
43) Isopropyl Acetate	8.688	43	37946m	5.625	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050225\
Data File : VN086462.D
Acq On : 02 May 2025 14:40
Operator : JC\MD
Sample : PB167816ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
PB167816ZHE#07

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/05/2025
Supervised By :Semsettin Yesilyurt 05/05/2025


Quant Time: May 02 23:16:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N041525W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 16 04:19:23 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40




Instrument :
MSVOA_N
ClientSampleId :
PB167816ZHE#07




Tgt Ion:168 Resp: 154085
Ion Ratio Lower Upper
168 100
99 59.9 52.5 78.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/05/2025
Supervised By :Semsettin Yesilyurt 05/05/2025




#16
Acetone
Concen: 27.562 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40




Tgt Ion: 43 Resp: 25215
Ion Ratio Lower Upper
43 100
58 28.2 25.3 37.9





#20
Methylene Chloride
Concen: 4.098 ug/l
RT: 5.283 min Scan# 51
Delta R.T. 0.012 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40




Instrument :
MSVOA_N
ClientSampleId :
PB167816ZHE#07



Tgt Ion: 84 Resp: 8464
Ion Ratio Lower Upper
84 100
49 107.5 98.2 147.2
51 32.9 29.8 44.6
86 61.8 52.0 78.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/05/2025
Supervised By :Semsettin Yesilyurt 05/05/2025



#33
1,2-Dichloroethane-d4
Concen: 50.331 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40

Tgt Ion: 65 Resp: 112459
Ion Ratio Lower Upper
65 100
67 52.3 0.0 106.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN086462.D

Acq: 02 May 2025 14:40

Instrument :

MSVOA_N

ClientSampleId :

PB167816ZHE#07

Tgt Ion:114 Resp: 29317





Ion	Ratio	Lower	Upper
114	100		
63	20.2	0.0	42.6
88	15.4	0.0	31.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/05/2025

Supervised By :Semsettin Yesilyurt 05/05/2025



#35

Dibromofluoromethane

Concen: 53.688 ug/l

RT: 8.165 min Scan# 1056




Delta R.T. -0.000 min


Lab File: VN086462.D

Acq: 02 May 2025 14:40

Tgt Ion:113 Resp: 73051


Ion	Ratio	Lower	Upper
113	100		
111	105.4	83.4	125.0
192	17.6	13.7	20.5





#43
Isopropyl Acetate
Concen: 5.625 ug/l m
RT: 8.688 min Scan# 1145
Delta R.T. 0.005 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40




Instrument :
MSVOA_N
ClientSampleId :
PB167816ZHE#07



Tgt Ion: 43 Resp: 37940
Ion Ratio Lower Upper
43 100
61 11.4 20.5 30.7
87 7.2 10.5 15.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/05/2025
Supervised By :Semsettin Yesilyurt 05/05/2025



#50
Toluene-d8
Concen: 51.288 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40


Tgt Ion: 98 Resp: 372970
Ion Ratio Lower Upper
98 100
100 65.9 52.5 78.7





#62
4-Bromofluorobenzene
Concen: 48.208 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40




Instrument :
MSVOA_N
ClientSampleId :
PB167816ZHE#07




Tgt Ion: 95 Resp: 127868
Ion Ratio Lower Upper
95 100
174 73.5 0.0 133.4
176 70.4 0.0 129.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/05/2025
Supervised By :Semsettin Yesilyurt 05/05/2025




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VN086462.D
Acq: 02 May 2025 14:40




Tgt Ion:117 Resp: 277547
Ion Ratio Lower Upper
117 100
82 53.5 44.7 67.1
119 31.8 26.4 39.6





#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.788 min Scan# 20
 Delta R.T. -0.000 min
 Lab File: VN086462.D
 Acq: 02 May 2025 14:40

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167816ZHE#07



Tgt Ion:152 Resp: 113320
 Ion Ratio Lower Upper
 152 100
 115 60.5 31.9 95.9
 150 155.2 0.0 352.0

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/05/2025
 Supervised By :Semsettin Yesilyurt 05/05/2025

