

Data File : VN055380.D
Acq On : 3 May 2019 11:12
Operator : JC/SP
Sample : K2658-08 10X
Misc : 5.17g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-E3-4(17-20)

Quant Time: May 03 23:53:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	424837	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	750113	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	690477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	298994	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	293599	52.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.66%	
35) Dibromofluoromethane	7.59	113	235254	47.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.56%	
50) Toluene-d8	10.09	98	946135	53.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.44%	
62) 4-Bromofluorobenzene	12.40	95	505058	90.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	181.80%	
Target Compounds						
39) Methylcyclohexane	9.09	83	18021	2.359	ug/l	95
73) Isopropylbenzene	12.25	105	683883	27.311	ug/l	99
78) n-propylbenzene	12.59	91	2051294	66.789	ug/l	100
83) tert-Butylbenzene	12.99	119	143893	6.118	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	34479	1.645	ug/l	99
85) sec-Butylbenzene	13.17	105	1129555	44.336	ug/l #	63
89) n-Butylbenzene	13.61	91	756672	50.853	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055380.D
Acq On : 3 May 2019 11:12
Operator : JC/SP
Sample : K2658-08 10X
Misc : 5.17g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-E3-4(17-20)

Quant Time: May 03 23:53:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

