

Data File : VN055388.D
Acq On : 3 May 2019 14:32
Operator : JC/SP
Sample : K2658-12DL 100X
Misc : 6.07g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-E2-3(17-20)DL

Quant Time: May 04 00:23:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	448492	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	781948	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	700879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	256606	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	309151	52.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.38%	
35) Dibromofluoromethane	7.59	113	243047	47.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.72%	
50) Toluene-d8	10.09	98	953366	51.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.88%	
62) 4-Bromofluorobenzene	12.40	95	328537	56.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.44%	

Target Compounds

						Qvalue
67) Ethyl Benzene	11.51	91	43583	1.761	ug/l	97
68) m/p-Xylenes	11.62	106	91654	10.377	ug/l	95
78) n-propylbenzene	12.59	91	225038	8.537	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	541758	26.930	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	1286748	71.519	ug/l	100
85) sec-Butylbenzene	13.17	105	132180	6.045	ug/l #	66
86) p-Isopropyltoluene	13.29	119	138226	7.691	ug/l	99
89) n-Butylbenzene	13.61	91	73837	5.782	ug/l #	78
95) Naphthalene	15.13	128	16986	5.551	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055388.D
Acq On : 3 May 2019 14:32
Operator : JC/SP
Sample : K2658-12DL 100X
Misc : 6.07g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-E2-3(17-20)DL

Quant Time: May 04 00:23:17 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

