

Data File : VN055391.D
Acq On : 3 May 2019 15:47
Operator : JC/SP
Sample : K2658-01DL 100X
Misc : 6.10g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-W2(14-17)DL

Quant Time: May 06 06:33:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	449220	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	766891	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	695199	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	245108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	304365	51.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.58%	
35) Dibromofluoromethane	7.59	113	248609	49.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.78%	
50) Toluene-d8	10.09	98	943829	51.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.86%	
62) 4-Bromofluorobenzene	12.40	95	310988	54.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.50%	
Target Compounds						
78) n-propylbenzene	12.59	91	174327	6.924	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	107190	5.578	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	667187	38.823	ug/l	100
85) sec-Butylbenzene	13.17	105	132485	6.343	ug/l #	78
86) p-Isopropyltoluene	13.29	119	123835	7.214	ug/l	100
89) n-Butylbenzene	13.61	91	100389	8.230	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055391.D
Acq On : 3 May 2019 15:47
Operator : JC/SP
Sample : K2658-01DL 100X
Misc : 6.10g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP1-SB-W2(14-17)DL

Quant Time: May 06 06:33:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

