

Data File : VN055396.D
Acq On : 3 May 2019 17:52
Operator : JC/SP
Sample : K2668-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: May 04 01:14:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	462632	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	791738	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	721954	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	224557	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	314821	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
35) Dibromofluoromethane	7.59	113	249901	48.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.18%	
50) Toluene-d8	10.09	98	971136	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	12.40	95	321464	54.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.62%	
Target Compounds						
16) Acetone	3.82	43	11960	6.342	ug/l	98
20) Methylene Chloride	4.55	84	821292	142.682	ug/l	99
43) Isopropyl Acetate	8.17	43	12890	1.380	ug/l #	89
95) Naphthalene	15.13	128	6404	4.286	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055396.D
Acq On : 3 May 2019 17:52
Operator : JC/SP
Sample : K2668-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: May 04 01:14:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

