

Data File : VN055398.D
Acq On : 3 May 2019 18:42
Operator : JC/SP
Sample : K2635-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-05-050219-C

Quant Time: May 04 01:19:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	472052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	805340	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	735530	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	243146	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	317382	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	7.59	113	257854	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	10.09	98	994195	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	12.40	95	331613	55.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.18%	
Target Compounds						
16) Acetone	3.83	43	16902	8.784	ug/l	99
20) Methylene Chloride	4.55	84	113864	19.387	ug/l	98
43) Isopropyl Acetate	8.17	43	13673	1.439	ug/l #	89
95) Naphthalene	15.13	128	21913	6.355	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055398.D
Acq On : 3 May 2019 18:42
Operator : JC/SP
Sample : K2635-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-05-050219-C

Quant Time: May 04 01:19:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

