

Data File : VN055400.D
Acq On : 3 May 2019 19:33
Operator : JC/SP
Sample : K2674-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FILTERBAG-RO

Quant Time: May 04 01:25:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	476678	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	827268	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	751707	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	229134	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	325854	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	7.59	113	261546	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
50) Toluene-d8	10.09	98	1014715	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	12.40	95	331057	54.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.06%	
Target Compounds						
16) Acetone	3.82	43	14101	7.257	ug/l	98
43) Isopropyl Acetate	8.17	43	15895	1.629	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055400.D
Acq On : 3 May 2019 19:33
Operator : JC/SP
Sample : K2674-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FILTERBAG-RO

Quant Time: May 04 01:25:18 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

