

Data File : VN055401.D
Acq On : 3 May 2019 19:58
Operator : JC/SP
Sample : K2668-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: May 04 01:29:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	485512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	847469	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	794570	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	251714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	325008	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	7.59	113	266869	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
50) Toluene-d8	10.09	98	1058481	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.40	95	349240	55.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.26%	
Target Compounds						
16) Acetone	3.82	43	17371	8.777	ug/l	89
20) Methylene Chloride	4.56	84	66651	11.034	ug/l	96
43) Isopropyl Acetate	8.18	43	19836	1.984	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VN055401.D
Acq On : 3 May 2019 19:58
Operator : JC/SP
Sample : K2668-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: May 04 01:29:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042919W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 30 07:16:52 2019
Response via : Initial Calibration

