

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050322\
 Data File : VN072312.D
 Acq On : 03 May 2022 18:33
 Operator : JC\MD
 Sample : N2619-26
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: May 04 05:14:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	179110	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	359345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	387954	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	150042	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	147975	53.814	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.620%
35) Dibromofluoromethane	8.022	113	112966	49.115	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.240%
50) Toluene-d8	10.439	98	461978	51.117	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.240%
62) 4-Bromofluorobenzene	12.727	95	173743	57.919	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050322\
Data File : VN072312.D
Acq On : 03 May 2022 18:33
Operator : JC\MD
Sample : N2619-26
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: May 04 05:14:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

