


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050322\
 Data File : VN072304.D
 Acq On : 03 May 2022 14:33
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Title : SW846 8260
 Last Update : Tue Apr 26 18:50:11 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1818

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	7739	PASS
75	95	30	60	56.0	20427	PASS
95	95	100	100	100.0	36507	PASS
96	95	5	9	7.5	2751	PASS
173	174	0.00	2	0.8	215	PASS
174	95	50	100	72.5	26453	PASS
175	174	5	9	7.8	2070	PASS
176	174	95	101	97.9	25893	PASS
177	176	5	9	5.6	1451	PASS