

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
 Data File : VN077555.D
 Acq On : 03 May 2023 12:06
 Operator : JC\MD
 Sample : PB152533TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152533TB

Quant Time: May 04 04:43:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	615038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1177236	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1102505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	387720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	411076	51.221	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.440%	
35) Dibromofluoromethane	8.172	113	360161	50.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.800%	
50) Toluene-d8	10.566	98	1444503	53.267	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.540%	
62) 4-Bromofluorobenzene	12.854	95	464182	52.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.640%	
Target Compounds						Qvalue
16) Acetone	4.443	43	50355	19.722	ug/l	94
20) Methylene Chloride	5.284	84	31154	3.848	ug/l	93
25) 2-Butanone	7.490	43	38376	10.597	ug/l	92
37) Ethyl Acetate	7.566	43	99258	12.778	ug/l	97
43) Isopropyl Acetate	8.690	43	410569	30.552	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
Data File : VN077555.D
Acq On : 03 May 2023 12:06
Operator : JC\MD
Sample : PB152533TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152533TB

Quant Time: May 04 04:43:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

