

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
 Data File : VN077557.D
 Acq On : 03 May 2023 12:54
 Operator : JC\MD
 Sample : PB152533ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152533ZHE#02

Quant Time: May 04 04:44:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	642442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1211335	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1107625	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	384851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	420417	50.150	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.300%	
35) Dibromofluoromethane	8.172	113	376804	51.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.500%	
50) Toluene-d8	10.566	98	1508347	54.055	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.120%	
62) 4-Bromofluorobenzene	12.854	95	471521	51.649	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.300%	
Target Compounds						
						Qvalue
16) Acetone	4.448	43	46683	17.504	ug/l	99
20) Methylene Chloride	5.289	84	31492	3.723	ug/l #	84
25) 2-Butanone	7.495	43	33029	8.732	ug/l	97
37) Ethyl Acetate	7.566	43	92232	11.539	ug/l	94
43) Isopropyl Acetate	8.689	43	436136	31.541	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
Data File : VN077557.D
Acq On : 03 May 2023 12:54
Operator : JC\MD
Sample : PB152533ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152533ZHE#02

Quant Time: May 04 04:44:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

