

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
 Data File : VN077559.D
 Acq On : 03 May 2023 13:43
 Operator : JC\MD
 Sample : PB152533ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152533ZHE#04

Quant Time: May 04 04:45:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	619517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1166767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	1104916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	383663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	409846	50.699	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	101.400%		
35) Dibromofluoromethane	8.172	113	356956	50.901	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.800%		
50) Toluene-d8	10.571	98	1470430	54.709	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	109.420%		
62) 4-Bromofluorobenzene	12.854	95	458791	52.174	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.340%		
Target Compounds						
						Qvalue
16) Acetone	4.442	43	45167	17.562	ug/l	97
20) Methylene Chloride	5.301	84	33634	4.124	ug/l #	79
25) 2-Butanone	7.495	43	30186	8.275	ug/l	92
37) Ethyl Acetate	7.572	43	94626	12.291	ug/l	97
43) Isopropyl Acetate	8.695	43	446634	33.534	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
Data File : VN077559.D
Acq On : 03 May 2023 13:43
Operator : JC\MD
Sample : PB152533ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152533ZHE#04

Quant Time: May 04 04:45:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

