

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
 Data File : VN077563.D
 Acq On : 03 May 2023 15:20
 Operator : JC\MD
 Sample : PB152533ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB152533ZHE#08

Quant Time: May 04 04:47:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	602366	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1160246	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1095644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	391374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	398011	50.636	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.280%
35) Dibromofluoromethane	8.172	113	354663	50.858	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	10.571	98	1455351	54.453	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.900%
62) 4-Bromofluorobenzene	12.854	95	458583	52.443	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.880%
Target Compounds						
						Qvalue
16) Acetone	4.437	43	48430	19.367	ug/l	93
20) Methylene Chloride	5.289	84	30090	3.794	ug/l	90
25) 2-Butanone	7.489	43	35769	10.085	ug/l	94
37) Ethyl Acetate	7.572	43	96606	12.619	ug/l	97
43) Isopropyl Acetate	8.695	43	489694	36.974	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
Data File : VN077563.D
Acq On : 03 May 2023 15:20
Operator : JC\MD
Sample : PB152533ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB152533ZHE#08

Quant Time: May 04 04:47:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

