

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
 Data File : VN077567.D
 Acq On : 03 May 2023 16:57
 Operator : JC\MD
 Sample : 02566-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1987

Quant Time: May 04 04:50:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	670843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1285649	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	1185594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	502371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	430123	49.136	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.280%		
35) Dibromofluoromethane	8.172	113	386082	49.963	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.920%		
50) Toluene-d8	10.566	98	1572968	53.113	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	106.220%		
62) 4-Bromofluorobenzene	12.848	95	537732	55.496	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	111.000%		
Target Compounds						
					Qvalue	
16) Acetone	4.437	43	206874	74.284	ug/l	89
20) Methylene Chloride	5.290	84	80376	9.101	ug/l #	86
25) 2-Butanone	7.495	43	23646	5.986	ug/l	97
37) Ethyl Acetate	7.566	43	97036	11.439	ug/l	97
40) Benzene	8.613	78	3263051	95.947	ug/l	99
43) Isopropyl Acetate	8.695	43	513194	34.969	ug/l #	85
48) Methyl methacrylate	9.660	41	17566	2.630	ug/l #	29
51) 4-Methyl-2-Pentanone	10.442	43	55689	6.469	ug/l	87
52) Toluene	10.630	92	2078049	93.272	ug/l	100
67) Ethyl Benzene	11.966	91	322227	7.296	ug/l	98
68) m/p-Xylenes	12.072	106	690138	40.207	ug/l	97
69) o-Xylene	12.401	106	660087	39.097	ug/l	100
80) 1,3,5-Trimethylbenzene	13.177	105	177281	5.121	ug/l	98
84) 1,2,4-Trimethylbenzene	13.489	105	670562	20.740	ug/l	90
95) Naphthalene	15.642	128	9031087	189.220	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
Data File : VN077567.D
Acq On : 03 May 2023 16:57
Operator : JC\MD
Sample : 02566-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1987

Quant Time: May 04 04:50:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

