

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
 Data File : VN077572.D
 Acq On : 03 May 2023 18:59
 Operator : JC\MD
 Sample : 02566-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1992

Quant Time: May 04 04:56:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	755438	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1463405	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	1343281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	532352	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	498777	50.598	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.200%
35) Dibromofluoromethane	8.172	113	450001	51.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.320%
50) Toluene-d8	10.571	98	1773945	52.623	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.240%
62) 4-Bromofluorobenzene	12.854	95	599522	54.358	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.720%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.531	59	11696	7.391	ug/l #	84
16) Acetone	4.437	43	210417	67.095	ug/l	98
20) Methylene Chloride	5.289	84	85168	8.563	ug/l #	85
25) 2-Butanone	7.495	43	23649	5.317	ug/l	90
37) Ethyl Acetate	7.566	43	98329	10.183	ug/l	95
40) Benzene	8.613	78	2390913	61.763	ug/l	99
43) Isopropyl Acetate	8.695	43	529981	31.726	ug/l #	84
52) Toluene	10.636	92	1530699	60.359	ug/l	100
67) Ethyl Benzene	11.966	91	238978	4.776	ug/l	98
68) m/p-Xylenes	12.071	106	499382	25.678	ug/l	100
69) o-Xylene	12.401	106	462482	24.177	ug/l	99
80) 1,3,5-Trimethylbenzene	13.177	105	118565	3.232	ug/l	95
84) 1,2,4-Trimethylbenzene	13.489	105	467726	13.652	ug/l	89
95) Naphthalene	15.642	128	6723459	153.933	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050323\
Data File : VN077572.D
Acq On : 03 May 2023 18:59
Operator : JC\MD
Sample : 02566-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1992

Quant Time: May 04 04:56:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

