

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN050324\  
 Data File : VN081996.D  
 Acq On : 03 May 2024 17:00  
 Operator : JC\MD  
 Sample : P2379-03 20X  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ORA-1226

Quant Time: May 06 01:57:55 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042524W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Apr 26 05:26:49 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 8.224  | 168   | 151591   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100  | 114   | 323112   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.865 | 117   | 290707   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.794 | 152   | 109624   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.583  | 65    | 125717   | 49.138   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =     | 98.280%  |
| 35) Dibromofluoromethane    | 8.165  | 113   | 93981    | 48.309   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 96.620%  |
| 50) Toluene-d8              | 10.565 | 98    | 402538   | 52.685   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =     | 105.360% |
| 62) 4-Bromofluorobenzene    | 12.847 | 95    | 141756   | 46.934   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 64 - 133 | Recovery | =     | 93.860%  |

Target Compounds Qvalue

-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN050324\  
Data File : VN081996.D  
Acq On : 03 May 2024 17:00  
Operator : JC\MD  
Sample : P2379-03 20X  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
ORA-1226

Quant Time: May 06 01:57:55 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042524W.M  
Quant Title : SW846 8260  
QLast Update : Fri Apr 26 05:26:49 2024  
Response via : Initial Calibration

