

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050420\
 Data File : VN061271.D
 Acq On : 4 May 2020 12:34
 Operator : JC/MD
 Sample : L2445-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK-4-29-2020

Quant Time: May 05 04:25:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	102181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	183664	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	174229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	72985	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	81939	50.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.36%	
35) Dibromofluoromethane	7.54	113	52575	48.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.82%	
50) Toluene-d8	10.06	98	227990	51.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.00%	
62) 4-Bromofluorobenzene	12.38	95	82823	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
Target Compounds						
16) Acetone	3.77	43	5701	3.674	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050420\
Data File : VN061271.D
Acq On : 4 May 2020 12:34
Operator : JC/MD
Sample : L2445-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK-4-29-2020

Quant Time: May 05 04:25:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

