

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050420\
 Data File : VN061272.D
 Acq On : 4 May 2020 12:57
 Operator : JC/MD
 Sample : PB128636TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB128636TB

Quant Time: May 05 04:28:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	105007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	190226	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	187953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	78415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	87242	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	7.55	113	50215	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
50) Toluene-d8	10.06	98	243048	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
62) 4-Bromofluorobenzene	12.38	95	88894	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
Target Compounds						
16) Acetone	3.78	43	11003	11.860	ug/l	94
20) Methylene Chloride	4.50	84	1729	1.335	ug/l	96
43) Isopropyl Acetate	8.13	43	40401	10.905	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050420\
Data File : VN061272.D
Acq On : 4 May 2020 12:57
Operator : JC/MD
Sample : PB128636TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB128636TB

Quant Time: May 05 04:28:35 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

