

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN050420\
 Data File : VN061273.D
 Acq On : 4 May 2020 13:20
 Operator : JC/MD
 Sample : L2467-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS040MW015-200428

Quant Time: May 05 04:37:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 29 08:14:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	105908	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	195005	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	186253	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	82091	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	85276	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
35) Dibromofluoromethane	7.55	113	56867	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	10.06	98	247183	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
62) 4-Bromofluorobenzene	12.37	95	92491	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
Target Compounds						
11) Tert butyl alcohol	4.73	59	7795	34.724	ug/l #	84
83) tert-Butylbenzene	12.97	119	7613	1.808	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN050420\
Data File : VN061273.D
Acq On : 4 May 2020 13:20
Operator : JC/MD
Sample : L2467-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS040MW015-200428

Quant Time: May 05 04:37:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 29 08:14:26 2020
Response via : Initial Calibration

