

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066937.D
 Acq On : 4 May 2021 15:50
 Operator : JC/MD
 Sample : PB136139TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139TB

Quant Time: May 05 05:24:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 13:11:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	547410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	907306	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	847669	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	305275	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	412580	47.88	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.76%
35) Dibromofluoromethane	7.514	113	298321	48.42	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.84%
50) Toluene-d8	10.030	98	1186958	50.59	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.18%
62) 4-Bromofluorobenzene	12.350	95	364907	45.93	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.86%
Target Compounds						
						Qvalue
16) Acetone	3.760	43	152921	50.45	ug/l	98
20) Methylene Chloride	4.473	84	292782	38.08	ug/l	92
43) Isopropyl Acetate	8.096	43	63178	4.04	ug/l #	89
95) Naphthalene	15.091	128	54873	4.09	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066937.D
Acq On : 4 May 2021 15:50
Operator : JC/MD
Sample : PB136139TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139TB

Quant Time: May 05 05:24:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Tue May 04 13:11:03 2021
Response via : Initial Calibration

