

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066941.D
 Acq On : 4 May 2021 17:28
 Operator : JC/MD
 Sample : PB136139ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#04

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:33:07 PM

Quant Time: May 05 05:25:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 13:11:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	531099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	890648	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	849688	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	309860	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.952	65	412962	49.40	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.80%		
35) Dibromofluoromethane	7.514	113	294038	48.62	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.24%		
50) Toluene-d8	10.030	98	1180844	51.27	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.54%		
62) 4-Bromofluorobenzene	12.350	95	380683	48.81	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	97.62%		
Target Compounds						Qvalue
16) Acetone	3.757	43	59662	20.29	ug/l	97
18) Methyl Acetate	4.264	43	12595m	1.50	ug/l	
20) Methylene Chloride	4.473	84	84894	11.38	ug/l #	90
43) Isopropyl Acetate	8.096	43	60923	3.97	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066941.D
Acq On : 4 May 2021 17:28
Operator : JC/MD
Sample : PB136139ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#04

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:33:07 PM

Quant Time: May 05 05:25:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Tue May 04 13:11:03 2021
Response via : Initial Calibration

