

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066944.D
 Acq On : 4 May 2021 18:41
 Operator : JC/MD
 Sample : PB136139ZHE#07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#07

Quant Time: May 05 05:26:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 13:11:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	499811	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	832489	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.349	117	804877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	298402	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	392119	49.84	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.68%
35) Dibromofluoromethane	7.512	113	283168	50.09	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.18%
50) Toluene-d8	10.030	98	1108443	51.49	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.98%
62) 4-Bromofluorobenzene	12.350	95	354424	48.62	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.24%
Target Compounds						
					Qvalue	
16) Acetone	3.757	43	157012	56.73	ug/l	95
18) Methyl Acetate	4.261	43	11538	1.46	ug/l	95
20) Methylene Chloride	4.468	84	277401	39.52	ug/l	93
43) Isopropyl Acetate	8.099	43	59795	4.17	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066944.D
Acq On : 4 May 2021 18:41
Operator : JC/MD
Sample : PB136139ZHE#07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#07

Quant Time: May 05 05:26:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Tue May 04 13:11:03 2021
Response via : Initial Calibration

