

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066945.D
 Acq On : 4 May 2021 19:05
 Operator : JC/MD
 Sample : PB136139ZHE#08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#08

Quant Time: May 05 05:27:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 13:11:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	491619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	822299	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	789103	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	275530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	392570	50.73	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.46%
35) Dibromofluoromethane	7.514	113	279250	50.01	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.02%
50) Toluene-d8	10.030	98	1105994	52.01	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.02%
62) 4-Bromofluorobenzene	12.350	95	351533	48.82	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.64%
Target Compounds						
					Qvalue	
16) Acetone	3.757	43	152009	55.84	ug/l	99
18) Methyl Acetate	4.261	43	10954	1.41	ug/l #	68
20) Methylene Chloride	4.467	84	280182	40.58	ug/l	93
43) Isopropyl Acetate	8.096	43	61694	4.35	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066945.D
Acq On : 4 May 2021 19:05
Operator : JC/MD
Sample : PB136139ZHE#08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#08

Quant Time: May 05 05:27:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Tue May 04 13:11:03 2021
Response via : Initial Calibration

