

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066946.D
 Acq On : 4 May 2021 19:29
 Operator : JC/MD
 Sample : PB136139ZHE#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#09

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:33:11 PM

Quant Time: May 05 05:27:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 13:11:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	506398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	844405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	819212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	298373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	394187	49.45	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.90%		
35) Dibromofluoromethane	7.512	113	279721	48.78	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.56%		
50) Toluene-d8	10.030	98	1124124	51.48	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.96%		
62) 4-Bromofluorobenzene	12.350	95	339559	45.92	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.84%		
Target Compounds					Qvalue	
16) Acetone	3.757	43	157749	56.26	ug/l	100
18) Methyl Acetate	4.269	43	12033m	1.50	ug/l	
20) Methylene Chloride	4.470	84	279016	39.23	ug/l	91
43) Isopropyl Acetate	8.096	43	61293	4.21	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066946.D
Acq On : 4 May 2021 19:29
Operator : JC/MD
Sample : PB136139ZHE#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#09

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:33:11 PM

Quant Time: May 05 05:27:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Tue May 04 13:11:03 2021
Response via : Initial Calibration

