

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066947.D
 Acq On : 4 May 2021 19:53
 Operator : JC/MD
 Sample : PB136139ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#10

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:33:13 PM

Quant Time: May 05 05:27:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 13:11:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	494618	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.520	114	836400	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	810852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	286016	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	401881	51.62	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.24%	
35) Dibromofluoromethane	7.512	113	283258	49.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.74%	
50) Toluene-d8	10.030	98	1124126	51.97	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.94%	
62) 4-Bromofluorobenzene	12.350	95	356370	48.66	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 97.32%	
Target Compounds						
16) Acetone	3.757	43	155422	56.75	ug/l	Qvalue 97
18) Methyl Acetate	4.272	43	14633m	1.87	ug/l	
20) Methylene Chloride	4.470	84	275279	39.63	ug/l	91
43) Isopropyl Acetate	8.096	43	59472	4.13	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066947.D
Acq On : 4 May 2021 19:53
Operator : JC/MD
Sample : PB136139ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#10

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:33:13 PM

Quant Time: May 05 05:27:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Tue May 04 13:11:03 2021
Response via : Initial Calibration

