

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066957.D
 Acq On : 5 May 2021 00:44
 Operator : JC/MD
 Sample : PB136139ZHE#14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#14

Quant Time: May 05 07:47:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	501737	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	849420	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	800308	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	295044	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	398253	50.43	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.86%			
35) Dibromofluoromethane	7.512	113	288368	49.99	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.98%			
50) Toluene-d8	10.030	98	1127793	51.34	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.68%			
62) 4-Bromofluorobenzene	12.350	95	346013	46.52	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 93.04%			
Target Compounds						
16) Acetone	3.760	43	35237	12.68	ug/l	96
18) Methyl Acetate	4.258	43	8483	1.07	ug/l #	72
20) Methylene Chloride	4.468	84	31348	4.45	ug/l	94
43) Isopropyl Acetate	8.096	43	62569	4.27	ug/l #	86
95) Naphthalene	15.094	128	36578	2.82	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066957.D
Acq On : 5 May 2021 00:44
Operator : JC/MD
Sample : PB136139ZHE#14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#14

Quant Time: May 05 07:47:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

