

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066958.D
 Acq On : 5 May 2021 1:08
 Operator : JC/MD
 Sample : PB136139ZHE#15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136139ZHE#15

Quant Time: May 05 07:48:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	488919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	823759	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	771193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	259415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	399142	51.87	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.74%	
35) Dibromofluoromethane	7.512	113	275636	49.28	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.56%	
50) Toluene-d8	10.030	98	1089880	51.16	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.32%	
62) 4-Bromofluorobenzene	12.350	95	335253	46.48	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.96%	
Target Compounds						
						Qvalue
16) Acetone	3.754	43	35785	13.22	ug/l	95
18) Methyl Acetate	4.258	43	9194	1.19	ug/l #	83
20) Methylene Chloride	4.468	84	33678	4.90	ug/l	95
43) Isopropyl Acetate	8.099	43	62866	4.43	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066958.D
Acq On : 5 May 2021 1:08
Operator : JC/MD
Sample : PB136139ZHE#15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136139ZHE#15

Quant Time: May 05 07:48:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

