

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066960.D
 Acq On : 5 May 2021 1:57
 Operator : JC/MD
 Sample : M2256-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R1-A-TOP

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:34:39 PM

Quant Time: May 05 07:48:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	470841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	795129	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	753626	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	267252	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	383430	51.74	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.48%	
35) Dibromofluoromethane	7.514	113	274384	50.82	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.64%	
50) Toluene-d8	10.030	98	1072707	52.17	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.34%	
62) 4-Bromofluorobenzene	12.350	95	323496	46.46	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.92%	
Target Compounds						
16) Acetone	3.757	43	61110	23.44	ug/l	Qvalue 97
18) Methyl Acetate	4.267	43	11401m	1.53	ug/l	
20) Methylene Chloride	4.470	84	62569	9.46	ug/l	91
43) Isopropyl Acetate	8.096	43	56408	4.12	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066960.D
Acq On : 5 May 2021 1:57
Operator : JC/MD
Sample : M2256-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R1-A-TOP

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:34:39 PM

Quant Time: May 05 07:48:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

