

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066967.D
 Acq On : 5 May 2021 4:45
 Operator : JC/MD
 Sample : M2256-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R1-D-BOTTOM

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:34:47 PM

Quant Time: May 05 07:50:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	459424	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	775964	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.352	117	737699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	253156	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	377528	52.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.42%			
35) Dibromofluoromethane	7.512	113	269794	51.20	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.40%			
50) Toluene-d8	10.030	98	1044321	52.04	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.08%			
62) 4-Bromofluorobenzene	12.350	95	301494	44.37	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 88.74%			
Target Compounds						Qvalue
16) Acetone	3.757	43	59074	23.22	ug/l	100
18) Methyl Acetate	4.258	43	8561m	1.18	ug/l	
20) Methylene Chloride	4.473	84	63228	9.80	ug/l	96
43) Isopropyl Acetate	8.096	43	52579	3.93	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066967.D
Acq On : 5 May 2021 4:45
Operator : JC/MD
Sample : M2256-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R1-D-BOTTOM

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:34:47 PM

Quant Time: May 05 07:50:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

