

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066968.D
 Acq On : 5 May 2021 5:09
 Operator : JC/MD
 Sample : M2256-15
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R2-A-TOP

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:34:49 PM

Quant Time: May 05 07:51:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	446031	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.517	114	765097	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	735624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	251903	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	379813	54.10	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 108.20%			
35) Dibromofluoromethane	7.514	113	267207	51.43	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.86%			
50) Toluene-d8	10.030	98	1034836	52.30	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.60%			
62) 4-Bromofluorobenzene	12.350	95	309045	46.13	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 92.26%			
Target Compounds						Qvalue
16) Acetone	3.754	43	53503	21.66	ug/l	98
18) Methyl Acetate	4.266	43	8482m	1.20	ug/l	
20) Methylene Chloride	4.470	84	60218	9.61	ug/l #	82
43) Isopropyl Acetate	8.099	43	52691	4.00	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066968.D
Acq On : 5 May 2021 5:09
Operator : JC/MD
Sample : M2256-15
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R2-A-TOP

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:34:49 PM

Quant Time: May 05 07:51:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

