

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066969.D
 Acq On : 5 May 2021 5:33
 Operator : JC/MD
 Sample : M2256-16
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R2-A-BOTTOM

Quant Time: May 05 07:51:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	456093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	769319	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	732991	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	244255	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	379159	52.81	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 105.62%	
35) Dibromofluoromethane	7.514	113	269124	51.52	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.04%	
50) Toluene-d8	10.030	98	1041903	52.37	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.74%	
62) 4-Bromofluorobenzene	12.350	95	309745	45.98	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.96%	
Target Compounds						
						Qvalue
16) Acetone	3.754	43	55596	22.01	ug/l	98
20) Methylene Chloride	4.470	84	63336	9.89	ug/l	94
43) Isopropyl Acetate	8.096	43	51861	3.91	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066969.D
Acq On : 5 May 2021 5:33
Operator : JC/MD
Sample : M2256-16
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R2-A-BOTTOM

Quant Time: May 05 07:51:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

