

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066972.D
 Acq On : 5 May 2021 6:45
 Operator : JC/MD
 Sample : M2256-19
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R2-C-TOP

Quant Time: May 05 07:52:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.589	168	454263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	771242	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	718555	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	241291	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	382196	53.45	ug/l	0.00
Spiked Amount 50.000	Range 61	- 141	Recovery	=	106.90%	
35) Dibromofluoromethane	7.514	113	266517	50.89	ug/l	0.00
Spiked Amount 50.000	Range 69	- 133	Recovery	=	101.78%	
50) Toluene-d8	10.030	98	1039750	52.13	ug/l	0.00
Spiked Amount 50.000	Range 65	- 126	Recovery	=	104.26%	
62) 4-Bromofluorobenzene	12.350	95	297681	44.08	ug/l	0.00
Spiked Amount 50.000	Range 58	- 135	Recovery	=	88.16%	
Target Compounds						
16) Acetone	3.754	43	67793	26.95	ug/l	98
20) Methylene Chloride	4.473	84	61651	9.66	ug/l	90
43) Isopropyl Acetate	8.094	43	48591	3.66	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066972.D
Acq On : 5 May 2021 6:45
Operator : JC/MD
Sample : M2256-19
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R2-C-TOP

Quant Time: May 05 07:52:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

