

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066973.D
 Acq On : 5 May 2021 7:09
 Operator : JC/MD
 Sample : M2256-20
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R2-C-BOTTOM

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:34:52 PM

Quant Time: May 05 07:52:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	467041	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	794190	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	754511	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	254155	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.957	65	387228	52.67	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.34%			
35) Dibromofluoromethane	7.514	113	276458	51.26	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.52%			
50) Toluene-d8	10.030	98	1074513	52.32	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.64%			
62) 4-Bromofluorobenzene	12.350	95	303052m	43.58	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 87.16%			
Target Compounds						Qvalue
16) Acetone	3.757	43	69523	26.88	ug/l	99
18) Methyl Acetate	4.256	43	9137m	1.24	ug/l	
20) Methylene Chloride	4.473	84	61885	9.43	ug/l	90
43) Isopropyl Acetate	8.099	43	51652	3.77	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066973.D
Acq On : 5 May 2021 7:09
Operator : JC/MD
Sample : M2256-20
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R2-C-BOTTOM

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:34:52 PM

Quant Time: May 05 07:52:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

