

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
 Data File : VN066975.D
 Acq On : 5 May 2021 7:58
 Operator : JC/MD
 Sample : M2256-22
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R2-D-BOTTOM

Manual Integrations
 APPROVED

MMDadoda
 5/5/2021 5:34:56 PM

Quant Time: May 05 08:27:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 07:24:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.592	168	468649	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.520	114	811149	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	759562	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.294	152	265188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.954	65	393875	53.39	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 106.78%			
35) Dibromofluoromethane	7.512	113	276909	50.27	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.54%			
50) Toluene-d8	10.030	98	1080103	51.49	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.98%			
62) 4-Bromofluorobenzene	12.350	95	295542	41.61	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 83.22%			
Target Compounds						Qvalue
16) Acetone	3.757	43	58292	22.46	ug/l	100
18) Methyl Acetate	4.256	43	9524m	1.28	ug/l	
20) Methylene Chloride	4.465	84	65622	9.97	ug/l	95
43) Isopropyl Acetate	8.096	43	54395	3.89	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050421\
Data File : VN066975.D
Acq On : 5 May 2021 7:58
Operator : JC/MD
Sample : M2256-22
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R2-D-BOTTOM

Manual Integrations
APPROVED

MMDadoda
5/5/2021 5:34:56 PM

Quant Time: May 05 08:27:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N050421W.M
Quant Title : SW846 8260
QLast Update : Wed May 05 07:24:52 2021
Response via : Initial Calibration

