


Ethyl methacrylate

Response Ratio



$$\text{Response} = 6.696\text{e-}001 * \text{Amt} - 6.951\text{e-}002$$


Coef of Det (r^2) = 0.996615 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N050421W.M

Calibration Table Last Updated: Wed May 05 07:24:52 2021

Styrene

Response Ratio



$$\text{Response} = 1.337\text{e}+000 * \text{Amt} - 1.260\text{e}-001$$


Coef of Det (r^2) = 0.997558 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N050421W.M

Calibration Table Last Updated: Wed May 05 07:24:52 2021

2-Hexanone

Response Ratio



$$\text{Response} = 4.105\text{e-}001 * \text{Amt} - 2.329\text{e-}001$$


Coef of Det (r^2) = 0.996776 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N050421W.M

Calibration Table Last Updated: Wed May 05 07:24:52 2021

n-Butylbenzene

Response Ratio




$$\text{Response} = 3.174\text{e}+000 * \text{Amt} - 3.326\text{e}-001$$

Coef of Det (r^2) = 0.996513 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N050421W.M

Calibration Table Last Updated: Wed May 05 07:24:52 2021

trans-1,4-Dichloro-2-butene



Response = $4.272 \times 10^{-1} \times \text{Amt} - 4.727 \times 10^{-2}$
Coef of Det (r^2) = 0.998544 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N050421W.M
Calibration Table Last Updated: Wed May 05 07:24:52 2021