

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
 Data File : VN072320.D
 Acq On : 04 May 2022 11:42
 Operator : JC\MD
 Sample : VN0504WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0504WBL01

Quant Time: May 04 14:19:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	214488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	415785	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	449072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	165667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	177242	53.826	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.660%	
35) Dibromofluoromethane	8.022	113	134277	50.456	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.920%	
50) Toluene-d8	10.439	98	545802	52.194	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.380%	
62) 4-Bromofluorobenzene	12.727	95	197996	57.044	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 114.080%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN050422\
Data File : VN072320.D
Acq On : 04 May 2022 11:42
Operator : JC\MD
Sample : VN0504WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0504WBL01

Quant Time: May 04 14:19:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

